

CERTIFICATE OF CALIBRATION

Date Of Issue 30 January 2017
Issue Number 1
Certificate Number TERUKAS45950
Issue By TER Calibration
Page 1 of 3



UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

0149



TER CALIBRATION LTD

Unit 1, Armstrong Point, Wigan, WN2 4AU
Tel:-01942 882275 Fax:-01942 897958
E:-me@ter.co.uk Web:-www.ter.co.uk

Approved Signatory
Les Finnen

Submitted By MED-LAB a Cytex
Company
Copeland Street
Derby
DE1 2PU
Engineer DAVEB
Procedure Number 010010153
Order Number PC35216
Date Received 26 January 2017
Calibration Date 29 January 2017
Request Recalibration 28 January 2018
Equipment RS 612-029 Tachometer
Serial Number 001772
Owners Identification
TERID 309686
JobNumber 658606
Conditions of Test
Temperature 20°C ±1°C
Humidity 43% ±10%

Method Of Test

The instrument was operated in accordance with the manufacturer's instruction manual. All results are recorded in tables 1 and 2.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Date Of Issue 30 January 2017
Issue Number 1
Certificate Number TERUKAS45950
Issue By TER Calibration
Page 2 of 3

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

Test 1 Pulse Output

Range	Applied RPM/Frequency	Output Frequency	Specification	Uncertainty of Measurement
100 - 6 000RPM	200.0 RPM(3.333 3Hz)	200.08 Hz	±1%	±0.03Hz
100 - 6 000RPM	1 000.0 RPM(16.666 6Hz)	1 000.03 Hz	±1%	±0.11Hz
100 - 6 000RPM	3 000.0 RPM(50.000 0Hz)	3 000.0 Hz	±1%	±0.3Hz
100 - 6 000RPM	6 000.0 RPM(100.000 0Hz)	5 999.9 Hz	±1%	±0.7Hz
1 000 - 6 000RPM	1 000.0 RPM(16.666 6Hz)	100.00 Hz	±1%	±0.02Hz
1 000 - 6 000RPM	5 000.0 RPM(83.333 3Hz)	499.99 Hz	±1%	±0.06Hz
1 000 - 6 000RPM	1 000.0 RPM(166.666Hz)	999.97 Hz	±1%	±0.11Hz
1 000 - 6 000RPM	3 000.0 RPM(500.000Hz)	2 999.9 Hz	±1%	±0.3Hz
1 000 - 6 000RPM	6 000.0 RPM(1 000.00Hz)	5 999.8 Hz	±1%	±0.7Hz

Test 2 DC Output

Range	Applied RPM/Frequency	Output Frequency	Specification	Uncertainty of Measurement
100 - 6 000RPM	200.0 RPM(3.333 3Hz)	200.1 mV	±1%	±0.1mV
100 - 6 000RPM	1 000.0 RPM(16.666 6Hz)	1 000.1 mV	±1%	±0.2mV
100 - 6 000RPM	3 000.0 RPM(50.000 0Hz)	2.998 V	±1%	±0.001V
100 - 6 000RPM	6 000.0 RPM(100.000 0Hz)	6.009 V	±1%	±0.002V
1 000 - 6 000RPM	1 000.0 RPM(16.666 6Hz)	99.74 mV	±1%	±0.01mV
1 000 - 6 000RPM	5 000.0 RPM(83.333 3Hz)	499.8 mV	±1%	±0.2mV
1 000 - 6 000RPM	1 000.0 RPM(166.666Hz)	0.999 8 V	±1%	±0.000 2V
1 000 - 6 000RPM	3 000.0 RPM(500.000Hz)	3.002 V	±1%	±0.001V
1 000 - 6 000RPM	6 000.0 RPM(1 000.00Hz)	6.009 V	±1%	±0.002V

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

CERTIFICATE OF CALIBRATION

Date Of Issue 30 January 2017
Issue Number 1
Certificate Number TERUKAS45950
Issue By TER Calibration
Page 3 of 3

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149
