

Med-Lab
 Derby, UK

CERTIFICATE OF CONFORMITY

Standardisation of Precision Conductivity Meter (D4308) Serial Number 16474 with KSLA Cell 48801

This certifies that the above instrument and accompanying cell have been examined and calibrated according to the requirements of the manufacturer. Performance has been assessed by checking the electrical conductivity of a kerosene fuel sample against that obtained using Precision Conductivity instrument, Serial No: 16474, retained for that purpose. Any differences between the measured values were found to be within the precision required by the ASTM D4308 procedure.

Please note that this certificate is valid only for the particular combination of meter and cell shown above. Any noticeable tampering or physical / electrical damage to the instrument or probe will render any calibration or warranty invalid.

Calibration Values

<u>Item</u>	<u>As Received</u>	<u>Post Calibration</u>	<u>Item</u>	<u>As Received</u>	<u>Post Calibration</u>
Zero@2000	0	0	502@2000	502	502
Zero@200	00.0	00.0	101@2000	101	101
Zero@20	0.01	0.00	101@200	101.3	101.0
CAL	1000	1000	10.00@20	9.62	10.00
Post Calibration					
Instrument Reading (16474)			365	R =	57
Reference Reading (107124)			367	r =	44
Reference Reading (119844)			367		

 Signed
 Date



 12, October 2017


Authorised Service Centre



Authorised Distributor