

# CERTIFICATE OF CALIBRATION

Date Of Issue 04 April 2023  
Issue Number 1  
Certificate Number TERUKAS77698  
Issue By TER Calibration  
Page 1 of 3



UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

0149



## TER CALIBRATION LTD

Unit 1, Armstrong Point, Wigan, WN2 4AU  
Tel:-01942 882275 Fax:-01942 897958  
E:-me@ter.co.uk Web:-www.ter.co.uk

Approved Signatory  
Les Finnen

A handwritten signature in black ink, appearing to read 'Les Finnen'.

Submitted By MED-LAB a Cytex  
Company  
Copeland Street  
Derby  
DE1 2PU  
Engineer THOMASH  
Procedure Number 010004187  
Order Number B2304-089  
Date Received 28 March 2023  
Calibration Date 03 April 2023  
Request Recalibration 02 April 2024  
Equipment FLUKE 73 SERIES 3 Digital Multimeter  
Serial Number 78640328  
Owners Identification  
TERID 325300  
JobNumber 742099  
Conditions of Test  
Temperature 20°C ±1°C  
Humidity 46% ±10%

### Method Of Test

The instrument was operated in accordance with the manufacturer's instruction manual. All results are recorded in tables 1 to 6.

The instrument was within specification at all points tested, with due allowance made for the uncertainty of measurement.

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor of  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to Expression of Uncertainty in Measurement and is inclusive of the unit under test. The uncertainties relate only to the measured values and do not carry any implication regarding the long term stability of the instrument.*

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

Date Of Issue 04 April 2023  
Issue Number 1  
Certificate Number TERUKAS77698  
Issue By TER Calibration  
Page 2 of 3

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

---

## Test 1 Direct Voltage

Range	Applied Voltage	Unit under Test	Specification	Uncertainty of Measurement
300 mV	300.00 mV	300.2 mV	$\pm(0.3\% + 1\text{dgt})$	0.1mV
3.2 V	2.700 0 V	2.702 V	$\pm(0.3\% + 1\text{dgt})$	0.001V
32 V	27.000 V	27.02 V	$\pm(0.3\% + 1\text{dgt})$	0.01V
32 V	20.000 V	20.01 V	$\pm(0.3\% + 1\text{dgt})$	0.01V
32 V	10.000 V	10.00 V	$\pm(0.3\% + 1\text{dgt})$	0.01V
32 V	1.000 V	1.00 V	$\pm(0.3\% + 1\text{dgt})$	0.01V
32 V	-27.000 V	-27.02 V	$\pm(0.3\% + 1\text{dgt})$	0.01V
320 V	270.00 V	270.2 V	$\pm(0.3\% + 1\text{dgt})$	0.1V
600 V	600.00 V	600 V	$\pm(0.4\% + 1\text{dgt})$	1V

## Test 2 Alternating Voltage

Range	Applied Voltage	Unit Under Test	Specification	Uncertainty of Measurement
3.2 V	2.700 0 V @ 100 Hz	2.701 V	$\pm(2\% + 2\text{dgts})$	0.004V
3.2 V	2.700 0 V @ 500 Hz	2.689 V	$\pm(2\% + 2\text{dgts})$	0.004V
32 V	27.000 V @ 100 Hz	27.02 V	$\pm(2\% + 2\text{dgts})$	0.04V
320 V	270.00 V @ 100 Hz	270.2 V	$\pm(2\% + 2\text{dgts})$	0.4V
600 V	500.00 V @ 100 Hz	500 V	$\pm(2\% + 2\text{dgts})$	2V
600 V	500.00 V @ 1 kHz	499 V	$\pm(2\% + 2\text{dgts})$	2V

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.*

# CERTIFICATE OF CALIBRATION

Date Of Issue 04 April 2023  
Issue Number 1  
Certificate Number TERUKAS77698  
Issue By TER Calibration  
Page 3 of 3

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

---

## Test 3 Resistance

Range	Applied Resistance	Unit Under Test	Specification	Uncertainty of Measurement
320 ohm	100.00 ohm	100.0 ohms	$\pm(0.5\% + 2\text{dgt})$	0.1ohm
3 200 kohm	1.000 0 kohm	1 000 ohms	$\pm(0.5\% + 1\text{dgt})$	0.001kohm
32 kohm	10.000 kohm	10.00 kohms	$\pm(0.5\% + 1\text{dgt})$	0.01kohm
320 kohm	100.00 kohm	100.0 kohms	$\pm(0.5\% + 1\text{dgt})$	0.1kohm
3.2 Mohm	1.000 0 Mohm	1.000 Mohms	$\pm(0.5\% + 1\text{dgt})$	0.001Mohm
32 Mohm	10.000 Mohm	10.02 Mohms	$\pm(2\% + 1\text{dgt})$	0.02Mohm

## Test 4 Direct Current

Range	Applied Current	Unit Under Test	Specification	Uncertainty of Measurement
32 mA	27.000 mA	26.94 mA	$\pm(1.5\% + 2\text{dgt})$	0.04mA
320 mA	270.00 mA	269.9 mA	$\pm(1.5\% + 2\text{dgt})$	0.4mA
10 A	10.000 A	10.00 A	$\pm(1.5\% + 2\text{dgt})$	0.02A

## Test 5 Alternating Current

Range	Applied Current	Unit Under Test	Specification	Uncertainty of Measurement
32 mA	27.000 mA @ 60 Hz	26.94 mA	$\pm(2.5\% + 2\text{dgt})$	0.05mA
320 mA	270.00 mA @ 60 Hz	269.9 mA	$\pm(2.5\% + 2\text{dgt})$	0.5mA
10 A	10.000 A @ 60 Hz	10.00 A	$\pm(2.5\% + 2\text{dgt})$	0.03A

## Test 6 Diode Test

Range	Applied Voltage	Unit Under Test	Specification	Uncertainty of Measurement
2 V	1.000 0 V	1.000 V	$\pm 2\%$ Typical	0.001V

The calibration was performed at the laboratory's permanent address.

The calibration relates only to the item listed on page 1.

END OF RESULTS

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.*