

CERTIFICATE OF CALIBRATION

Date Of Issue 30 January 2017
Issue Number 1
Certificate Number TERUKAS45949
Issue By TER Calibration
Page 1 of 2



UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

0149



TER CALIBRATION LTD

Unit 1, Armstrong Point, Wigan, WN2 4AU
Tel:-01942 882275 Fax:-01942 897958
E:-me@ter.co.uk Web:-www.ter.co.uk

Approved Signatory
Les Finnen

Submitted By MED-LAB a Cytex
Company
Copeland Street
Derby
DE1 2PU
Engineer DAVEB
Procedure Number 010007967
Order Number PC35216
Date Received 25 January 2017
Calibration Date 28 January 2017
Request Recalibration 27 January 2018
Equipment COMPACT CT7 Tachometer
Serial Number 714000
Owners Identification
TERID 309685
JobNumber 658605
Conditions of Test
Temperature 20°C ±1°C
Humidity 43% ±10%

Method Of Test

The instrument was operated in accordance with the manufacturer's instruction manual. All results are recorded in table 1.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to the units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Date Of Issue 30 January 2017
Issue Number 1
Certificate Number TERUKAS45949
Issue By TER Calibration
Page 2 of 2

UKAS ACCREDITED CALIBRATION LABORATORY NO. 0149

Test 1 Revs per minute

Range	Applied Frequency	Calculated RPM	Unit Under Test	Specification	Uncertainty of Measurement
3 - 99 999RPM	0.500 00 Hz	30.0 RPM	30 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	0.833 33 Hz	50.0 RPM	50 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	1.666 66 Hz	100.0 RPM	100 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	3.333 33 Hz	200.0 RPM	200 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	5.000 00 Hz	300.0 RPM	300 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	8.333 33 Hz	500.0 RPM	500 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	16.666 6 Hz	1 000.0 RPM	1 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	33.333 3 Hz	2 000.0 RPM	2 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	50.000 0 Hz	3 000.0 RPM	3 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	83.333 3 Hz	5 000.0 RPM	5 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	166.666 6 Hz	10 000.0 RPM	10 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	333.333 3 Hz	20 000.0 RPM	20 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	500.000 0 Hz	30 000.0 RPM	30 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	666.666 Hz	40 000.0 RPM	40 000 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	833.333 Hz	50 000.0 RPM	50 001 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	1 500.000 Hz	90 000.0 RPM	90 002 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$
3 - 99 999RPM	1 650.000 Hz	99 000.0 RPM	99 004 RPM	$\pm(0.05\% + 1\text{digit})$	$\pm(0.001\% + 1\text{RPM})$

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.